

Parameter	Value
Data File Name	// ktclab/ rymrdata/ openuseB/ data/ liyanyao/ nmr/ 3973-0294_P/ 26/ fid
Title	3973-0294_P.26.fid
Comment	ry_kwa2i c-13 (night, 2.5 hr, ns=5120) p31 CDCl3 / opt/ nmrdata liyanyao 17
Origin	Bruker BioSpin GmbH & Co. KG
Owner	nmrsu
Site	
Instrument	Avance
Author	
Solvent	CDCl3
Temperature	300.3
Pulse Sequence	zgpg30
Experiment	1D
Probe	Z167419_0066 (PI HR-BBO500S2- BBF/ H/ D-5.0-Z SP)
Number of Scans	16
Receiver Gain	101.0
Relaxation Delay	1.0000
Pulse Width	12.0000
Presaturation Frequency	
Acquisition Time	1.0486
Acquisition Date	2025-06-02T15:56:01
Modification Date	2025-06-02T15:56:34
Class	
Spectrometer Frequency	202.36
Spectral Width	62500.0
Lowest Frequency	-11015.7
Nucleus	31P
Acquired Size	65536
Spectral Size	131072
Digital Resolution	0.48

--16.05

